

# Detection and Discrimination of Influenza A, B, and Respiratory Syncytial Viruses using the 3M Integrated Cyclor

Michael Aye\*, Jules Chen, Lakshmi Nair, Yuan Xie, Huong Mai, Shannon Dempsey, Yin-Peng Chen, and Maurice Exner  
Focus Diagnostics Inc., Cypress, CA

## Abstract

**Objective:** Recently, it has become evident that many rapid tests designed to detect respiratory viruses lack sensitivity. As a result, a clinical need has arisen for more sensitive tests that will detect and discriminate influenza viruses. Molecular techniques have been shown to be more sensitive than culture or rapid lateral flow methodologies. Therefore, a multiplex RT-PCR assay was designed to detect Influenza A viruses (of all hemagglutinin and neuraminidase types), Influenza B viruses and Respiratory Syncytial virus (RSV).

**Methods:** Primers were designed to target conserved regions of Influenza A, B, and RSV. Specificity of the assay was tested by running a panel of known influenza strains and against a set of clinical specimens that were verified to be positive for the target viruses. In addition, a panel of other respiratory specimens was tested to determine whether there was any cross reactivity with these organisms. Assay sensitivity was determined analytically by limit of detection studies, and sensitivity was compared to culture, DFA, and RT-PCR methods.

**Results:** The assay was able to detect multiple influenza types and strains including seasonal H1 and H3 strains, H1N1 (2009), and H5N1. All influenza B strains tested were detected, as were all strains of RSV-A and RSV-B that were tested. The assay was shown to be more sensitive than culture and DFA for all viruses.

**Conclusions:** The use of this assay on the 3M integrated cyclor provides a sensitive, rapid and high throughput method for detecting and discriminating Influenza A, B, and RSV viruses. The portability and ease of use of the instrument make it useful as a first line tool for diagnosis of respiratory infections.

## Methods

**Virus strains.** The following virus strains were tested: **Influenza A:** Influenza A/PR/8/34 H1N1, 10<sup>9.5</sup> TCID<sub>50</sub>/mL (Advanced Biotechnologies, Inc., Columbia, MD), Influenza A/Hong Kong/8/68 H3N2, 10<sup>7.5</sup> TCID<sub>50</sub>/mL (ABI). **Influenza B:** Influenza B/Malaysia/2506/2004, 5.16x10<sup>8</sup> TCID<sub>50</sub>/mL (Virapur, San Diego, CA), Influenza B/Great Lakes/1739/54, 1.58x10<sup>7</sup> TCID<sub>50</sub>/mL (ATCC, Manassas, VA), **RSV:** RSV A1, 3.2x10<sup>6</sup> TCID<sub>50</sub>/mL (Virapur), RSV B2, 2.8x10<sup>6</sup> TCID<sub>50</sub>/mL (ATCC).

**Sample preparation.** 200 µL of each clinical specimen was extracted using the Roche MagNA Pure LC automated system with a Total Nucleic Acid Isolation Kit (Roche Diagnostics, Indianapolis, IN), and eluted in 50 µL of elution buffer. The same clinical specimens (200 µL of sample input) were extracted using a NucleiSens easyMAG (bioMérieux, Durham, NC), and eluted with 50 µL. An RNA internal control (Focus Diagnostics, Inc.) was added to each specimen and control prior to extraction to monitor the extraction process and to detect PCR inhibition.

**Primers.** Primers were designed to target a conserved region of the Influenza A matrix segment, Influenza B matrix segment and RSV M gene. An additional set of primers was used to detect the RNA internal control. Each probe was labeled with a different fluorescent dye with a distinct emission profile to distinguish the fluorescent signal from each target or internal control.

**Real-time PCR amplification and detection.** A one-step RT-PCR assay was carried out with 5 µL of extracted RNA and 5 µL of reaction mix (Simplexa™ Flu A/B & RSV assay, Focus Diagnostics, Inc.). RT-PCR was performed by using an Integrated Cyclor microfluidic platform (3M, St. Paul, MN) programmed with the following parameters: 1 cycle of 47°C for 10 min; 1 cycle of 97°C for 2 min; 45 cycles of 97°C for 5 sec and 58°C for 30 sec.

## Methods (Cont.)

Fluorescent signal for target-specific amplification was detected at 58°C. **Analytical specificity:** Genomic DNA or RNA of a variety of viral (≥10<sup>5</sup> TCID<sub>50</sub>/mL) and bacterial (≥10<sup>6</sup> CFU/mL) pathogens, or clinical specimens with Ct ≤ 30 for each targeted pathogen were tested to verify lack of cross-reactivity of the Simplexa™ assay with nucleic acids from other organisms. For each contrived sample, organisms were spiked into a 200 µL aliquot of Universal Transport Medium (UTM) (Diagnostic Hybrids, Athens, OH).

**Analytical sensitivity:** Influenza A, Influenza B and RSV strains were serially diluted (in UTM) and extracted in ten replicates. Each extracted sample was tested in duplicate wells for a limit of detection (LoD) study. Extracted nucleic acid samples from a defined TCID<sub>50</sub>/mL concentration of each virus strain were tested with the Simplexa™ assay.

**Clinical specimens:** 239 clinical specimens (Table 1) were extracted with two automated systems described above, and the Simplexa™ assay results were compared with previous DFA, culture or RT-PCR results. These specimens included nasal swabs, nasopharyngeal swabs, and swabs with unspecified respiratory sources. To eliminate bias, previously reported results were blinded to operators. Simplexa™ results were compared to previous results for concordance. Any discrepant specimens were tested by the Focus Diagnostics reference laboratory for additional resolution using Flu A/B or RSV RT-PCR testing.

Table 1. Clinical Specimen Summary

Specimen Type	Culture	DFA	RT-PCR	Total
Flu A Positive	18	21	0	39
Flu B Positive	11	23	22	56
Flu A/B Negative	0	0	51	51
RSV Positive	0	0	55	55
RSV Negative	0	0	38	38
<b>Total # of Clinical Specimens</b>				<b>239</b>

## Results


**Specificity Studies:** The Simplexa™ assay did not detect any of the pathogens listed in Table 2 and was specific for the targeted viruses. In addition, database searches indicated that the targeted regions did not have significant homology with sequences from other pathogens.

**Limit of Detection:** The LoD, defined as the lowest concentration with ≥95% detection, was determined to be <5 TCID<sub>50</sub>/mL for Influenza A, <30 TCID<sub>50</sub>/mL for influenza B, and <1 TCID<sub>50</sub>/mL for RSV (Table 3).

**Method Comparison Studies:** As shown in Tables 4-6, the Simplexa™ assay provided results with >95% positive and negative agreements compared to previous clinical results for Influenza A, B and RSV.

## Results (Cont.)

Figure 1. Examples of Amplification Curves with Simplexa™ Assay



Flu A - FAM signal (green), Flu B - JOE signal (blue), RSV - CFB610 signal (red), RNA Internal Control - Quasar 670 signal (purple)

Table 2. Specificity testing with different respiratory pathogens

Pathogen	Pathogen	Pathogen	Pathogen
Adenovirus 1	Enterovirus 71	Moraxella catarrhalis	Parainfluenza virus type-3
Adenovirus 7A	Epstein Barr Virus	Mumps	Pseudomonas aeruginosa
Bordetella pertussis	Escherichia coli O157H7	Mycobacterium tuberculosis	Rhinovirus A1
Chlamydia pneumoniae	Haemophilus influenzae	Mycoplasma pneumoniae	Staphylococcus aureus
Coronavirus 229E	Lactobacillus plantarum	Neisseria elongata	Staphylococcus epidermidis
Coronavirus OC43	Legionella longbeachae	Neisseria meningitidis	Streptococcus pneumoniae
Corynebacterium diphtheriae	Measles	Parainfluenza virus type-1	Streptococcus pyogenes
Cytomegalovirus	Human Metapneumovirus	Parainfluenza virus type-2	Streptococcus salivarius

Table 3. Analytical Sensitivity, Limit of Detection studies

Virus	Extraction Method	TCID <sub>50</sub> /mL	Average Ct (FAM)		Min Ct	Max Ct	Replicates Detected
			Mean	SD			
Influenza A	easyMAG	3.16	35.52	35.00	36.40	39.20	20/20
	MagNA Pure	3.16	37.22	35.80	42.40	39.20	20/20
	easyMAG	0.63	34.43	33.60	35.40	20/20	20/20
Influenza B	easyMAG	0.63	33.62	30.00	35.80	20/20	20/20
	MagNA Pure	25.8	34.95	34.20	35.60	20/20	20/20
	MagNA Pure	25.8	35.25	34.20	36.50	20/20	20/20
Influenza B/Malaysia/2506/2004	easyMAG	15.8	35.04	35.00	37.2	19/20	20/20
	MagNA Pure	15.8	35.94	35.10	36.2	20/20	20/20
	MagNA Pure	15.8	35.94	35.10	36.2	20/20	20/20
RSV	easyMAG	0.32	34.83	33.60	35.30	20/20	20/20
	MagNA Pure	0.32	35.27	34.50	36.90	20/20	20/20
	easyMAG	0.70	33.03	30.00	35.10	20/20	20/20
RSV B1	easyMAG	0.70	33.72	32.10	35.60	20/20	20/20
	MagNA Pure	0.70	33.72	32.10	35.60	20/20	20/20

## Results (Cont.)

Table 4. Simplexa™ Concordance for Flu A

Simplexa™ Flu A/B & RSV Real-Time RT-PCR	Previous Results			% Agreement
	Flu A Positive	Flu A Negative	Total	
	Flu A Positive	38	2*	
Flu A Negative	1**	198	199	% Negative Agreement: 99.0% (198/200) 95% CI: 96.4-99.7
<b>Total</b>	<b>39</b>	<b>200</b>	<b>239</b>	

\*Samples 41 and 75 were detected as Flu A positive by Simplexa™ assay, whereas previous PCR result detected these samples as Flu A negative. Simplexa™ results were identical for both samples by both extraction methods. Samples 41 and 75 were confirmed to be Flu A positive by Focus reference laboratory.

\*\* Sample 111 was detected as Flu A negative by Simplexa™ assay, whereas previous DFA result detected as Flu A positive. Simplexa™ results were identical for both extraction methods. Sample 111 was confirmed to be Flu A positive by Focus reference laboratory.

Table 5. Simplexa™ Concordance for Flu B

Simplexa™ Flu A/B & RSV Real-Time RT-PCR	Previous Results			% Agreement
	Flu B Positive	Flu B Negative	Total	
	Flu B Positive	55	1*	
Flu B Negative	1**	182	183	% Negative Agreement: 99.5% (182/183) 95% CI: 97.0-99.9
<b>Total</b>	<b>56</b>	<b>183</b>	<b>239</b>	

\*Sample 209 was detected as Flu B positive by the Simplexa™ assay, whereas the previous PCR result detected as Flu B negative. Simplexa™ results were identical for both extraction methods. Focus reference laboratory detected sample 209 as Flu B negative.

\*\* Sample 112 was detected as Flu B negative by the Simplexa™ assay, whereas previous culture results determined the sample to be Flu B positive. Simplexa™ results were identical for both extraction methods. Sample 112 was confirmed to be Flu B negative by Focus reference laboratory.

Table 6. Simplexa™ Concordance for RSV

Simplexa™ Flu A/B & RSV Real-Time RT-PCR	Previous Results			% Agreement
	RSV Positive	RSV Negative	Total	
	RSV Positive	55	1*	
RSV Negative	0	183	183	% Negative Agreement: 99.3% (183/184) 95% CI: 97.0-99.9
<b>Total</b>	<b>55</b>	<b>184</b>	<b>239</b>	

\*Sample 25 was detected as Flu A and RSV dual positive by the Simplexa™ assay, whereas previous culture results detected only Flu A. Simplexa™ results were identical for both extraction methods. Sample 25 was confirmed to be Flu A and RSV dual positive by Focus reference laboratory.

## Conclusions

- The Simplexa™ assay's limit of detection (LoD) for Influenza A, Influenza B, and RSV was <5 TCID<sub>50</sub>/mL, <30 TCID<sub>50</sub>/mL, and <1 TCID<sub>50</sub>/mL, respectively, with both automated extraction systems.
- Positive and negative agreement with previous clinical results demonstrated that the Simplexa™ Flu A/B & RSV assay could effectively detect seasonal influenza A, B and RSV.
- The Simplexa™ assay, using the 3M Integrated Cyclor, provides robust and sensitive detection for Influenza A, B, and RSV, and the features of the instrument (small footprint, high-throughput and rapid cycling) provides an effective option for clinical laboratory testing.

